
Technical Specifications
Equinox²
PC-Based Audiometer



Included and Optional Parts

AC440

Included parts:

- Equinox2.0 AC440 CD
- OtoAccess™ database CD
- TDH39 Audiometric headset or DD45 Audiometric headset
- MTH400 Headset
- EMS400 Talk back microphone
- B81 Bone conductor
- APS3 Patient response button
- Standard USB cable
- Power cable 120 or 230V
- Mouse pad
- Instructions for Use document

Optional parts:

- DAK70 Audiometer keyboard with live voice mic.
- Earphone 3A insert earphones (5As may be substituted)
- IP30 insert earphones
- B81 Bone Conductor
- B71 Bone Conductor
- ACC60 Equinox2.0 carrying case
- CIR22 Insert masking earphones
- Audiocup enclosures
- Peltor noise excluding headset
- HDA280 Audiometric headset
- HDA300 Audiometric headset
- KOSS R80 high frequency headset
- AP70 Power amplifier 2x70 Watt
- SP90 Loudspeaker
- SP85A Loudspeaker
- SP90A Loudspeaker
- AFC8 Sound cabin installation panel
- Optical USB 1.1 isolation extension cable

Optional special tests:

- High Frequency audiometry (HF440)
- Multi Frequency module (MF440)
- Speech from hard-drive (SFH440)
- SIS1 test
- Master Hearing Aid (MHA440),
- Hearing Loss Simulator (HLS440)
- Loudness Scaling (LS440)
- QuickSIN
- TEN test

General Technical Specifications

Equinox2.0 Hardware - Technical Specifications

Medical CE-mark:	The CE-mark indicates that Interacoustics A/S meets the requirements of Annex II of the Medical Device Directive 93/42/EEC Approval of the quality system is made by TÜV – identification no. 0123.	
Safety Standards	IEC 60601-1, UL60601-1, CAN/CSA-C22.2 No.60601-1 Class I, Applied parts type B, Continuous operation	
EMC Standard	IEC 60601-1-2 IEC 60645-1	
Calibration	Technical information is located in the specifications for the software modules. Calibration information and instructions are located in the Service manual.	
PC requirements:	1.6 GHz Dual core CPU or better (Intel recommended) 1GB RAM or more. (Windows 8: 1GB 32-bit; 2GB 64-bit) Hard drive with min 20 GB of free space. Minimum display resolution of 1024x768 pixels (1280x1024 or higher recommended) DirectX 9 graphics with WDDM 1.0 or higher. (Intel/Nvidia recommended) One or more USB ports, version 1.1 or higher. DVD-Rom drive.	
Operative System:	Windows XP or Windows Vista, Windows 7, Windows 8	
Display:	Minimum resolution of 1024x768 with hardware accelerated DirectX/Direct3D graphics card.	
Disc Space:	1GB RAM or more. (Windows 8: 1GB 32-bit; 2GB 64-bit)	
Compatible software	Noah 3.7, Noah 4., OtoAccess™ and XML compatible Affinity ^{2.0} / Equinox ^{2.0} Suite VSP, HLS, MHA (simulators)	
Input Specifications	Talk Back	330 μ Vrms at max. input gain for 0dB VU-reading Input impedance : 47.5K Ω
	Mic. 1/TF & Mic. 2	
	Pat. Resp. L & R	Switches 3.3V to the logic input. (The switch current is 33 μ A)
	Inp. Aux. 1 & 2	20mVrms at max. input gain for 0dB VU-reading Input impedance : 15K Ω
	TB Coupler	
	TB Coupler - internal TB (Equinox2.0⁰ only)	
	Insitu L & R - Probe mic.	
	CD1 & CD2	10mVrms at max input gain for 0dB VU-reading Input impedance : 10k Ω
	TB Ref.	7mVrms at max. input gain for 0dB VU-reading Input impedance : 4,3K Ω
	TB Ref – internal TB (Equinox2.0⁰ only)	
	Insitu L & R - Ref. mic	
	Ref.Mic./Ext.	Not in use
	Coupler/Ext.	
	Wave files	Plays wave file from hard disk drive
Output Specifications	FF1 / FF2 (Terminal Block)	Up to 12.6Vrms by 8 Ω load 70Hz-20kHz \pm 3dB
	TB Lsp.	
	FF1/ FF2	Up to 7Vrms by 600 Ω load 70Hz-20kHz \pm 3dB
	Sp 1, Sp 2, Sp 3, Sp 4	
	Left, Right	Up to 7.0Vrms by 10 Ω load 70Hz-20kHz \pm 3dB
	Ins. Left, Ins. Right	
	Bone	
	Ins. Mask.	
	HF/HLS	
	Insitu L, Insitu R	
	Monitor, Ass. Mon.	Max.3.5Vrms. by 8 Ω load 70Hz-20kHz \pm 3dB
Sp. 1-4 Power Out		

	DC	Voltage: 5VDC Current: 0.5A
	TB Loop	Up to 100mA/meter 70Hz-20kHz ±3dB
	FF Loop	
	Batt. Sim.	Voltage: 1.1 – 1.6VDC Impedance range: 0 – 25 Ω.
	Batt. Sim. - Internal TB (Equinox2.0⁰ only)	
Data Connections	USB/PC	USB B socket for connection to PC (compatible with USB 1.1 and later)
	USB	USB A socket for connection of other USB devices (Internal USB 1.1 hub)
	Keyb.	Serial Peripheral Interface Bus (SPI interface) Check the Service manual for more information.
Internal test box:	Built in test box holds telecoil drive as well as special dual speaker set for checking directional microphone function.	
Supported Systems	Windows [®] XP (SP2 or later and compatible), Windows [®] VISTA Windows [®] 7 (32 and 64 bit) Windows [®] 8 (32 and 64 bit)	
Dimensions (LxWxH)	Equinox2.0 ⁰ :	42 x 38 x 14 cm / 16.5 x 15 x 5.5 inches
Weight	Equinox2.0 ⁰ :	5.5 kg / 12.1 lbs.
Power supply	100-240 V~, 50-60Hz	
Power Consumption:	195VA	
Operation environment	Temperature:	15-35°C
	Re. Humidity:	30-90% Non condensing
Transport and storage	Transport temperature:	-20-50°C
	Storage temperature:	0-50°C
	Re. Humidity:	10-95% Non condensing

Technical Specifications of the AC440 Software

Medical CE-mark:	The CE-mark indicates that Interacoustics A/S meets the requirements of Annex II of the Medical Device Directive 93/42/EEC. Approval of the quality system is made by TÜV – identification no. 0123.	
Audiometer Standards:	Tone: IEC60645-1/ANSI S3.6 Type 1 Speech: IEC60645-2/ANSI S3.6 Type A or A-E	
Transducers & Calibration:	Calibration information and instructions are located in the Service manual. Check the accompanying Appendix for RETSPL levels for transducers	
Air Conduction		
DD45	PTB/DTU report 2009	Headband Static Force 4.5N ±0.5N
TDH39	ISO 389-1 1998, ANSI S3.6-2010	Headband Static Force 4.5N ±0.5N
HDA300	ISO 389-8 2006, ANSI S3.6-2010	Headband Static Force 8,8N ±0.5N
HDA280	PTB report 2004	Headband Static Force 5N ±0.5N
E.A.R Tone 3A/5A	ISO 389-2 1994, ANSI S3.6-2010	
IP30	ISO 389-2 1994, ANSI S3.6-2010 DES-2361	
CIR 33	ISO 389-2	
Bone Conduction	Placemenet: Mastoid	
B71	ISO 389-3 1994, ANSI S3.6-2010	Headband Static Force 5.4N ±0.5N
B81	ISO 389-3 1994, ANSI S3.6-2010	Headband Static Force 5.4N ±0.5N
Free Field	ISO 389-7 2005, ANSI S3.6-2010	
High Frequency	ISO 389-5 2004, ANSI S3.6-2010	
Effective masking	ISO 389-4 1994, ANSI S3.6-2010	
Patient Response switch:	Hand held push button.	
Patient communication:	Talk Forward and Talk Back.	
Monitor:	Output through external earphone or speaker.	
Stimuli:	Pure tone, Warble tone, NB, SN, WN, TEN noise	
Tone	125-20000Hz separated in two ranges 125-8000Hz and 8000-20000Hz. Resolution 1/2-1/24 octave.	
Warble Tone	1-10 Hz sine +/- 5% modulation	
Wave file	44100Hz sampling, 16 bits, 2 channels	
Masking	Automatic selection of narrow band noise (or white noise) for tone presentation and speech noise for speech presentation.	
Narrow band noise:	IEC 60645-1:2001, 5/12 Octave filter with the same centre frequency resolution as pure Tone.	
White noise:	80-20000Hz measured with constant bandwidth	
Speech Noise.	IEC 60645-2:1993 125-6000Hz falling 12dB/octave above 1KHz +/-5dB	
Presentation	Manual or Reverse. Single or multiple pulses. pulse time adjustable from 200mS-5000mS in 50mS steps. Simultaneous or alternating.	
Intensity	Check the accompanying Appendix for maximum output levels	
Steps	Available Intensity Steps is 1, 2 or 5dB	
Accuracy	Sound pressure levels: ± 2 dB. Vibration force levels: ± 5 dB.	
Extended range function	If not activated, the Air Conduction output will be limited to 20 dB below maximum output.	
Frequency	Range: 125Hz to 8kHz (Optional High Frequency: 8 kHz to 20 kHz) Accuracy: Better than ± 1 %	
Distortion (THD)	Sound pressure levels: below 1.5 % Vibration force levels: below 3 %.	
Signal Indicator(VU)	Time weighting: 350mS Dynamic range: -20dB to +3dB Rectifier characteristics: RMS Selectable inputs are provide with an attenuator by which the level can be adjusted to the indicator reference position(0dB)	
Storing capability:	Tone audiogram: dB HL, MCL, UCL, Tinnitus, R+L Speech Audiogram: WR1, WR2, WR3, MCL, UCL, Aided, Unaided, Binaural, R+L.	
Compatible Software:	Noah 4, Noah 3.7, OtoAccess™ and XML compatible	